ES-Module ES



GS Test Module for function tests of ECG devices

- ☑ ECG impulse output
- ☑ periodic signals
- ☑ arrhythmia output
- ☑ pacemaker outputs
- ☑ simulation of respiration
- ☑ signal output in accordance with IEC 60601

S.P.L. Elektronik

Mess- und Prüftechnik
für medizinische und industrielle Anwendungen
Tel. +49 33232 39915, www.spl-elektronik.com
Forstweg 1, 14656 Brieselang, Deutschland

TÜVRheinland ZERTIFIZIERT
Management System ISO 9001:2015
www.tuv.com ID 0910095181

Technical Data

ECG impulse amplitude:	1–5 mV ± 1% in 1 mV steps
ECG signal duration:	1–200 ms ± 1% in 1 ms steps
ECG impulse form:	sinus sinus square triangle rectangle trapeze ISO ventricular fibrillation ventricular tachycardia mains frequency NSR
ECG impulse frequency:	10 - 300 bpm ± 2 % in 1 bpm steps
Signal frequency variable:	1 - 100 Hz ± 2 % in 1 Hz steps
Signal frequency sine:	0,3 Hz ± 2 %
Respiration:	Base value ± 1 Ohm
Accessories:	10 x STA8 ECG Adapter-Clip 1 x STA30 ECG plug adapter 1 x STA32 ECG adapter line

Description of functions:

The GS test module ES serves as test-signal generator for ECG impulses. These ECG impulses can be used for the functional test of the signal-representation and signal-evaluation of ECG monitors. Furthermore, the extended ECG functionality can be tested over the respiration-function with apnoea-alarm-control.

The GS-X base unit and a PC are required for testing with this plug-in module. The execution of a signal-output can be done with the PC-Software. The results of the signal-representation can be determined, assessed and stored.

PC software can be used for a flexible output of the offered waveforms. Consequently, an integration into complex test instructions and into automatic test-sequences is possible. For that reason, a high-quality documentation of the signal-representation and signal-evaluation of an ECG appliance is possible.

Because of the, as far as possible, freely configurable and variable ECG signal forms, a large variety of signal sequences can be output. The simulation of arrhythmic signal forms offers the possibility to test more complex devices in accordance with a standard.

The generation of the calibration impulse CAL in accordance with IEC 60601-2-51 enables practical operating function-controls for manufacturers and examiners.

Level ratios:

Ratio	Voltage
Abl. I (L - R)	+ 1,00 mV
Abl. II (F - R)	+ 1,56 mV
Abl. III (F - L)	- 0,56 mV
N – R	+ 0,59 mV
N – L	+ 1,59 mV
N – F	+ 2,15 mV
N – C1	+ 0,59 mV
N – C2	+ 1,18 mV
N – C3	+ 1,75 mV
N – C4	+ 2,03 mV
N – C5	+ 2,83 mV
N – C6	+ 3,35 mV

(Technical modifications and errors reserved. 10/2026)

S.P.L. Elektronik

Mess- und Prüftechnik
für medizinische und industrielle Anwendungen
Tel. +49 33232 39915, www.spl-elektronik.com
Forstweg 1, 14656 Brieselang, Deutschland

TÜVRheinland ZERTIFIZIERT
Management System ISO 9001:2015
www.tuv.com ID 0910095181