Safety Tester **GM-410**



Measurement and test device for testing electrical safety in accordance with IEC 60601 / IEC 62353 / IEC 61010 / EN 50678 / EN 50699 MPBetreibV / BetrSichV / DGUV Vorschrift 3



- ☑ PC operation
- ☑ automatic or multimeter measurement
- ☑ robust light metal case
- ☑ 25 A PE measure in accordance with IEC 60601
- ☑ perfect for mobile working with laptop

S.P.L. Elektronik

Mess- und Prüftechnik
für medizinische und industrielle Anwendungen
Tel. +49 33232 39915, www.spl-elektronik.com
Forstweg 1, 14656 Brieselang, Deutschland

TÜVRheinland ZERTIFIZIERT
Management System ISO 9001:2015
www.tuv.com ID 0910095181

Technical Data

Line voltage:	230 V / 115 V ac, ± 10 %, 50/60 Hz
Output power:	maximum 3,5 kW
Protection class:	1
Overvoltage category:	II
Environmental temperature:	+ 5 - + 40 °C
Storage temperature:	- 10 - + 50 °C
Measurement range	
Voltages measurement:	0 - 300 V ac (input resistance: 10 MOhm)
Discrimination:	0,3 V
Earth conductor resistance:	0,00 - 40 Ohm (test voltage 6 V ac, max. 25 A / max. 5 A)
Discrimination:	10 MOhm
Insulation resistance:	0,2 - 100 MOhm (test voltage 500 V dc, max. 3,5 mA)
Discrimination:	0,1 – 2 MOhm
Leakage current:	0 - 10 mA or 0 – 20 mA
Resistance:	1000 Ohm ± 1 % or 2000 Ohm ± 1 %
Discrimination:	1 µA or 0,2 µA
Differential current:	10 µA - 20000 µA
Discrimination:	1 µA or 0,2 µA
Output power:	1 – 3,5 kW
Discrimination:	1 W
Current:	0 – 16 A
Discrimination:	10 mA

Intrinsic uncertainty

Measurement	Range	Error
Voltage:	0 - 300 V ac	± 0,3 V or ± 1 % of measurement value
Earth conductor resistance:	0,00 – 4,9 Ohm 5 – 40 Ohm	± 0,03 Ohm or ± 5 % of measurement value
Insulation resistance:	0,2 – 4,9 MOhm 5 – 100 MOhm	± 0,2 MOhm or ± 5 % of measurement value
Leakage current:	0 – 99 µA 100 – 20000 µA	± 2 µA or ± 1 % of measurement value
Differential current:	10 – 20000 µA	± 2 µA or ± 1 % of measurement value
Output power:	1 – 3,5 kW	± 2 W or ± 5 % of measurement value
Current:	0 – 16 A	± 50 mA or ± 2,5 % of measurement value

The specified intrinsic uncertainties relate to the respective measuring circuit. The operating uncertainty at the test object connections is ± 5%. The displayed value is normalized according to the documentation / standard requirement, if required.

Interface:	1 x USB for PC connection 1 x RS-232 for PC-connection 1 x RS-232 for further test appliances
Test object connection:	1 x outlet CEE7/4 12 x safety sockets (4mm) for applied part, groupable into 3 groups 1 x safety socket 4mm for test probe 1 x safety socket 4mm for TP3 1 x safety socket 4mm for PE 1 x safety socket 4mm for FE
Accessories:	1 x measurement line with test probe, 1 m length 1 x test adapter PA-X for self-diagnosis test 1 x USB cable 1 x power cord 16 A
Mechanical data:	light weight metal case IP20, 343 x 88 x 280 mm (W x H x D), approx. 6 kg

GM-410 is a measurement and test device for testing the electrical safety of medical technical and other technical appliances. The measurements and tests correspond to the conditions of IEC 60601, IEC 62353, IEC 61010, EN 50678 and EN 50699.

To control the GM-410 a 100% compatible IBM computer with industrial standards is needed. The communication between PC and GM-410 follows through the RS-232 serial interface and/or an USB interface. The control will be available among other things for the software such as

ACTIMED, WAVE, NOVA.

With GM-410 the following tests are realized:

- line voltage measurement
- output current measurement
- output power measurement
- leakage current measurement
- equivalent leakage current measurement
- earth conductor resistance measurement
- isolation resistance measurement
- differential current measurement

(Technical modifications and errors reserved. 10/2026)

S.P.L. Elektronik

Mess- und Prüftechnik
für medizinische und industrielle Anwendungen
Tel. +49 33232 39915, www.spl-elektronik.com
Forstweg 1, 14656 Brieselang, Deutschland

TÜVRheinland ZERTIFIZIERT
Management System ISO 9001:2015
www.tuv.com ID 0910095181